



Ref. Certif. No.

DE 2-032828

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME**CB TEST CERTIFICATE**

Product	All-In-One Multimeter
Name and address of the applicant	Pokit Innovations PTY.LTD. Suite 2.2, 56 Delhi Rd, North Ryde, Sydney NSW 2113, Australia
Name and address of the manufacturer	MediDiagno (Dongguan) Tech CO., Ltd Room 301, Building 8, No.15, Baiyunqian East Second Street, Dalang Town, Dongguan City, 523770 Guangdong, P.R. China
Name and address of the factory	MediDiagno (Dongguan) Tech CO., Ltd Room 301, Building 8, No.15, Baiyunqian East Second Street, Dalang Town, Dongguan City, 523770 Guangdong, P.R. China
Ratings and principal characteristics	3.7V, 130mAh, 0.481Wh (supplied by internal Li-ion battery) Measurement range: 600V CAT II, 10A (without tip shroud) 600V CAT III, 10A (with tip shroud)
Trademark (if any)	pokit PRO
Customer's Testing Facility (CTF) Stage used	N/A
Model / Type Ref.	POK-PRO
Additional information (if necessary may also be reported on page 2)	-see also test report CN21XWYS 001. Additionally evaluated according to IEC 61010-031.
A sample of the product was tested and found to be in conformity with	IEC 61010-1:2010+A1 IEC 61010-2-033:2012 for national differences see test report
As shown in the Test Report Ref. No. which forms part of this Certificate	CN21XWYS 001

This CB Test Certificate is issued by the National Certification Body

TÜV Rheinland LGA Products GmbH
Tillystr. 2, 90431 Nürnberg, Germany
Phone + 49 221 806-1371
Fax + 49 221 806-3935
Mail: cert-validity@de.tuv.com
Web : www.tuv.com

Date: 2021-11-01

Signature:

Dipl.-Ing. Univ. S. O. Steinke